
Midmark M11 - v1.0.0
Total Cycles: 1351

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2021 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.3 F 0.0
2:00 108.1 F 0.0
4:00 126.6 F 0.0
6:00 157.0 F 0.1
8:00 183.2 F 2.1
10:00 203.4 F 5.3
12:00 216.0 F 9.0
14:00 230.2 F 12.9
16:00 230.4 F 7.6
18:00 240.2 F 12.3
20:00 236.6 F 9.2
22:00 240.8 F 10.1
24:00 246.1 F 13.1
26:00 257.6 F 19.3
28:00 268.3 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.8 F 29.0
2:00 272.0 F 28.3
3:00 271.9 F 28.5
4:00 272.7 F 28.9

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 213.0 F 0.0
10:00 222.6 F 0.0
15:00 228.3 F 0.0
20:00 229.1 F 0.0
25:00 227.4 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 29:19
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:07:34
CYCLE COMPLETE

11/10/2021 11:07 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____