

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 004711  
16:38 14/11/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:21  
PRESSURIZING 2:24  
STERILIZING 3:16  
136.3 C 223kPa 3:16  
136.1 C 222kPa 3:46  
136.3 C 224kPa 4:16  
136.4 C 224kPa 4:46  
136.4 C 225kPa 5:16  
136.4 C 224kPa 5:46  
136.4 C 222kPa 6:16  
136.4 C 223kPa 6:46  
136.4 C 223kPa 6:46  
VENTING 6:46  
AIR DRYING 7:33  
CYCLE COMPLETE 12:38

-----  
Digital Signature #  
C517A9EACAD0BADB  
-----