
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001086
10:45 12/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:33
PRESSURIZING 3:34
STERILIZING 4:37
136.4 C 223kPa 4:37
136.2 C 222kPa 5:07
136.4 C 221kPa 5:37
136.1 C 217kPa 6:07
136.1 C 221kPa 6:37
136.4 C 223kPa 7:07
136.2 C 224kPa 7:37
136.4 C 223kPa 8:07
136.4 C 223kPa 8:07
VENTING 8:07
AIR DRYING 8:59
CYCLE COMPLETE 9:03

Digital Signature #
794D4E6EECCE810C
