

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R750  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.7uS / 0.4ppm  
CYCLE NUMBER 000843  
7:26 20/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:27  
PRESSURIZING 3:29  
STERILIZING 4:36  
136.5 C 223kPa 4:36  
136.4 C 223kPa 5:06  
135.9 C 219kPa 5:36  
136.4 C 223kPa 6:06  
136.0 C 220kPa 6:36  
136.4 C 221kPa 7:06  
136.4 C 221kPa 7:36  
136.2 C 223kPa 8:06  
136.2 C 223kPa 8:06  
VENTING 8:07  
AIR DRYING 9:00  
CYCLE COMPLETE 9:07

-----  
Digital Signature #  
23260C0EB663732D  
-----