

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4379

Name: RITTER2  
Sterilizer ID: V2313228  
02/25/2025 03:46 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 93.2 F 0.0  
2:00 94.5 F 0.5  
4:00 102.8 F 1.9  
6:00 132.0 F 4.5  
8:00 165.0 F 7.6  
10:00 190.8 F 10.9  
12:00 216.7 F 4.9  
14:00 221.6 F 7.5  
16:00 230.0 F 11.6  
18:00 230.3 F 7.6  
20:00 242.1 F 12.8  
22:00 240.5 F 11.2  
24:00 252.1 F 17.0  
26:00 262.4 F 23.1  
28:00 272.2 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.9  
1:00 271.8 F 30.2  
2:00 271.9 F 30.3  
3:00 271.8 F 30.2  
4:00 271.9 F 30.1

Min 271.6 F 29.8  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.6 F 0.0  
5:00 199.3 F 0.0  
10:00 203.3 F 0.0  
15:00 196.6 F 0.0  
20:00 192.3 F 0.0  
25:00 193.7 F 0.0  
30:00 194.2 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 29:10  
STERILIZING: 4:00  
VENTING: 3:52  
DRYING: 30:00  
TOTAL CYCLE: 01:08:14  
CYCLE COMPLETE

02/25/2025 04:54 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_