

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2783

Name: RITTER2  
Sterilizer ID: V2313228  
08/17/2023 10:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.6 F 0.0  
2:00 183.2 F 0.0  
4:00 209.6 F 2.3  
6:00 224.5 F 7.5  
8:00 236.4 F 12.9  
10:00 237.8 F 10.6  
12:00 236.2 F 9.3  
14:00 238.5 F 9.5  
16:00 251.0 F 16.1  
18:00 263.1 F 23.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.7  
1:00 272.6 F 29.5  
2:00 272.0 F 28.7  
3:00 272.2 F 29.5  
4:00 272.1 F 29.0  
  
Min 271.3 F 28.5  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
5:00 207.1 F 0.0  
10:00 200.5 F 0.0  
15:00 200.5 F 0.0  
20:00 197.6 F 0.0  
25:00 194.6 F 0.0  
30:00 191.6 F 0.0  
Drying Complete

FILLING: 1:43  
HEATING: 19:60  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:59:01  
CYCLE COMPLETE

08/17/2023 11:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_