
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001187
10:22 22/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:06
PRESSURIZING 2:14
STERILIZING 3:26
136.0 C 220kPa 3:26
136.4 C 224kPa 3:56
136.4 C 220kPa 4:26
136.3 C 222kPa 4:56
136.2 C 223kPa 5:26
136.1 C 222kPa 5:56
136.4 C 225kPa 6:26
136.4 C 221kPa 6:56
136.4 C 221kPa 6:56
VENTING 6:56
AIR DRYING 7:46
CYCLE COMPLETE 22:46

Digital Signature #
D00DAE9872145556
