
Midmark M11 - v1.0.0
Total Cycles: 696

Name: MIDMARK
Sterilizer ID: V2178198
02/24/2021 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 100.4 F 0.0
2:00 125.2 F 0.0
4:00 115.8 F 0.1
6:00 151.3 F 0.8
8:00 192.5 F 4.5
10:00 209.5 F 8.4
12:00 223.4 F 12.8
14:00 232.7 F 8.1
16:00 239.1 F 13.4
18:00 241.1 F 10.8
20:00 236.6 F 8.8
22:00 251.2 F 15.6
24:00 263.1 F 22.5
26:00 271.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 271.8 F 28.5
2:00 271.6 F 28.5
3:00 271.5 F 28.4
4:00 272.0 F 28.6

Min 271.4 F 28.4
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 177.8 F 0.0
5:00 241.4 F 0.0
10:00 228.6 F 0.0
15:00 232.3 F 0.0
20:00 228.1 F 0.0
25:00 193.7 F 0.0
30:00 194.9 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 26:45
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:04:29
CYCLE COMPLETE

02/24/2021 11:02 AM
ACCEPTED: X
REJECTED:
Operator: U01