
Midmark M11 - v1.0.5
Total Cycles: 569

Name: RITTER2
Sterilizer ID: V2313228
11/23/2021 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.0 F 0.0
2:00 82.4 F 0.0
4:00 112.9 F 0.0
6:00 114.5 F 0.1
8:00 145.4 F 0.2
10:00 173.2 F 0.4
12:00 195.3 F 1.6
14:00 209.8 F 4.5
16:00 220.5 F 7.8
18:00 227.3 F 11.4
20:00 227.9 F 6.6
22:00 239.2 F 11.7
24:00 237.5 F 9.7
26:00 239.2 F 9.8
28:00 249.0 F 15.0
30:00 260.2 F 21.1
32:00 270.3 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.5
2:00 271.9 F 28.7
3:00 271.6 F 28.9
4:00 272.7 F 29.4

Min 271.3 F 28.4
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.8 F 0.0
5:00 185.1 F 0.0
10:00 184.7 F 0.0
15:00 186.8 F 0.0
20:00 191.2 F 0.0
25:00 191.8 F 0.0
30:00 194.4 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 32:57
STERILIZING: 4:00
VENTING: 3:21
DRYING: 30:00
TOTAL CYCLE: 01:11:10
CYCLE COMPLETE

11/23/2021 01:30 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____