
Midmark M11 - v1.0.0
Total Cycles: 1520

Name: RITTER 1
Sterilizer ID: V2178198
01/13/2022 07:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.6 F 0.0
2:00 77.6 F 0.0
4:00 92.8 F 0.0
6:00 128.9 F 1.5
8:00 161.1 F 4.1
10:00 192.9 F 7.1
12:00 207.5 F 10.2
14:00 215.1 F 8.9
16:00 221.0 F 5.6
18:00 229.4 F 9.3
20:00 238.4 F 13.4
22:00 235.8 F 9.4
24:00 240.6 F 10.3
26:00 244.6 F 12.5
28:00 255.9 F 18.4
30:00 266.1 F 24.6
32:00 272.5 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.6
1:00 272.2 F 28.7
2:00 272.9 F 29.2
3:00 271.5 F 28.7
4:00 272.3 F 28.8

Min 271.5 F 28.4
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.1 F 0.0
5:00 231.0 F 0.0
10:00 190.4 F 0.0
15:00 184.9 F 0.0
20:00 184.3 F 0.0
25:00 186.0 F 0.0
30:00 192.4 F 0.0

Drying Complete

FILLING: 0:58
HEATING: 32:24
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:10:34
CYCLE COMPLETE

01/13/2022 09:28 AM
ACCEPTED: X
REJECTED:
Operator: U02