
Midmark M11 - v1.0.0
Total Cycles: 4474

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2024 03:56 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.7 F 0.0
2:00 119.1 F 1.0
4:00 124.0 F 3.4
6:00 155.2 F 6.6
8:00 176.0 F 10.6
10:00 203.0 F 4.0
12:00 214.4 F 6.8
14:00 225.6 F 10.7
16:00 229.9 F 7.4
18:00 238.9 F 11.4
20:00 236.3 F 9.0
22:00 249.3 F 15.1
24:00 260.4 F 21.3
26:00 270.6 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.6 F 29.1
2:00 271.4 F 28.4
3:00 272.3 F 29.3
4:00 272.2 F 28.8

Min 271.3 F 28.4
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 208.1 F 0.0
10:00 206.8 F 0.0
15:00 203.2 F 0.0
20:00 206.2 F 0.0
25:00 207.1 F 0.0
30:00 206.9 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 26:50
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:05:22
CYCLE COMPLETE

04/18/2024 05:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____