
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007239
10:58 22/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:02
PRESSURIZING 3:03
STERILIZING 4:02
136.0 C 220kPa 4:02
136.2 C 223kPa 4:32
136.3 C 225kPa 5:02
136.4 C 225kPa 5:32
136.3 C 222kPa 6:02
136.3 C 225kPa 6:32
136.4 C 221kPa 7:02
136.5 C 225kPa 7:32
136.5 C 225kPa 7:32
VENTING 7:32
AIR DRYING 8:27
CYCLE COMPLETE 38:27

Digital Signature #
90355F25455E163A
