
Midmark M11 - v1.0.0
Total Cycles: 1442

Name: RITTER 1
Sterilizer ID: V2178198
12/09/2021 12:36 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.9 F 0.0
2:00 167.8 F 3.3
4:00 185.6 F 7.4
6:00 206.7 F 12.6
8:00 223.0 F 7.0
10:00 236.5 F 12.3
12:00 236.7 F 9.6
14:00 237.1 F 8.7
16:00 249.3 F 14.8
18:00 261.8 F 21.8
20:00 272.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.3
1:00 272.5 F 28.7
2:00 272.0 F 28.6
3:00 272.3 F 28.6
4:00 271.9 F 28.5

Min 271.5 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 208.6 F 0.0
10:00 214.3 F 0.0
15:00 214.9 F 0.0
20:00 218.0 F 0.0
25:00 218.0 F 0.0
30:00 216.8 F 0.0
Drying Complete

FILLING: 1:56
HEATING: 20:20
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:59:23
CYCLE COMPLETE

12/09/2021 01:36 PM
ACCEPTED: X
REJECTED:
Operator: U02