
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001267
8:59 22/09/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:41
PRESSURIZING 4:41
STERILIZING 6:10
136.0 C 220kPa 6:10
136.4 C 223kPa 6:40
136.1 C 222kPa 7:10
136.3 C 222kPa 7:40
136.4 C 225kPa 8:10
136.3 C 223kPa 8:40
136.3 C 223kPa 9:10
136.4 C 225kPa 9:40
136.4 C 225kPa 9:40
VENTING 9:40
AIR DRYING 10:38
CYCLE COMPLETE 22:58

Digital Signature #
6ECC1CFBF7345A6A
