
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001588
16:32 13/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:32
STERILIZING 3:31
136.0 C 220kPa 3:31
136.3 C 224kPa 4:01
136.1 C 223kPa 4:31
136.4 C 226kPa 5:01
136.4 C 225kPa 5:31
136.2 C 223kPa 6:01
136.4 C 225kPa 6:31
136.4 C 225kPa 7:01
136.4 C 225kPa 7:01
VENTING 7:01
AIR DRYING 7:51
CYCLE COMPLETE 7:57

Digital Signature #
A115A39FF6666DA2
