

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
2.4uS / 1.5ppm  
CYCLE NUMBER 003680  
9:55 25/05/2022

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:15  
PRESSURIZING 2:18  
STERILIZING 3:19  
136.0 C 220kPa 3:19  
136.4 C 221kPa 3:49  
136.4 C 224kPa 4:19  
136.4 C 222kPa 4:49  
136.0 C 220kPa 5:19  
136.4 C 225kPa 5:49  
136.4 C 224kPa 6:19  
136.4 C 223kPa 6:49  
136.4 C 223kPa 6:49  
VENTING 6:49  
AIR DRYING 7:36  
CYCLE COMPLETE 8:10  
-----

Digital Signature #  
7E9D623D98AA19F3  
-----