
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005356
13:46 11/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:45
PRESSURIZING 1:51
STERILIZING 2:44
136.4 C 223kPa 2:44
136.4 C 221kPa 3:14
136.4 C 222kPa 3:44
136.4 C 217kPa 4:14
136.4 C 224kPa 4:44
136.4 C 225kPa 5:14
136.4 C 223kPa 5:44
136.3 C 219kPa 6:14
136.3 C 219kPa 6:14
VENTING 6:14
AIR DRYING 7:05
CYCLE COMPLETE 11:25

Digital Signature #
D66E1B2735F4C8D8
