
Midmark M11 - v1.0.0
Total Cycles: 5271

Name: RITTER 1
Sterilizer ID: V2178198
01/23/2025 06:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.3 F 0.0
2:00 170.1 F 3.6
4:00 180.8 F 7.7
6:00 201.1 F 12.4
8:00 222.2 F 5.7
10:00 231.2 F 10.7
12:00 231.6 F 6.9
14:00 243.9 F 12.5
16:00 240.5 F 10.7
18:00 253.4 F 17.0
20:00 264.4 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.4
1:00 272.2 F 29.0
2:00 271.2 F 28.3
3:00 272.1 F 29.2
4:00 272.5 F 29.3

Min 271.0 F 28.3
Max 273.2 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.3 F 0.0
5:00 214.1 F 0.0
10:00 199.2 F 0.0
15:00 201.4 F 0.0
20:00 210.8 F 0.0
25:00 238.1 F 0.0
30:00 237.3 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:57
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

01/23/2025 07:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____