
Midmark M11 - v1.0.5
Total Cycles: 3533

Name: RITTER2
Sterilizer ID: V2313228
04/16/2024 12:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 122.8 F 0.0
2:00 139.1 F 1.5
4:00 152.2 F 4.6
6:00 176.6 F 8.1
8:00 198.5 F 12.2
10:00 210.7 F 4.3
12:00 221.8 F 8.2
14:00 232.4 F 12.4
16:00 234.4 F 9.3
18:00 239.3 F 10.3
20:00 246.6 F 13.9
22:00 258.3 F 20.3
24:00 268.9 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.7 F 29.6
2:00 271.5 F 28.8
3:00 272.0 F 29.6
4:00 272.3 F 29.4

Min 271.2 F 28.8
Max 272.8 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.6 F 0.0
5:00 201.0 F 0.0
10:00 203.1 F 0.0
15:00 192.5 F 0.0
20:00 187.2 F 0.0
25:00 180.5 F 0.0
30:00 176.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 25:09
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:03:41
CYCLE COMPLETE

04/16/2024 01:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____