
Midmark M11 - v1.0.0
Total Cycles: 3792

Name: RITTER 1
Sterilizer ID: V2178198
09/18/2023 04:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.1 F 0.0
2:00 153.7 F 0.0
4:00 179.5 F 3.1
6:00 202.0 F 6.9
8:00 217.5 F 11.2
10:00 222.5 F 5.8
12:00 234.0 F 10.7
14:00 232.9 F 7.7
16:00 244.1 F 12.3
18:00 245.2 F 12.8
20:00 257.4 F 19.3
22:00 268.3 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 272.4 F 28.7
2:00 272.2 F 28.8
3:00 272.1 F 28.5
4:00 272.5 F 29.0

Min 271.6 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.5 F 0.0
5:00 214.8 F 0.0
10:00 235.8 F 0.0
15:00 232.9 F 0.0
20:00 225.2 F 0.0
25:00 223.1 F 0.0
30:00 217.8 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:12
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:01:43
CYCLE COMPLETE

09/18/2023 05:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____