
Midmark M11 - v1.0.0
Total Cycles: 2633

Name: RITTER 1
Sterilizer ID: V2178198
12/02/2022 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 76.4 F 0.0
2:00 79.8 F 0.0
4:00 90.0 F 0.1
6:00 120.2 F 2.2
8:00 149.2 F 5.4
10:00 174.6 F 9.5
12:00 193.9 F 7.6
14:00 217.3 F 6.0
16:00 228.9 F 10.1
18:00 235.9 F 8.6
20:00 240.8 F 11.2
22:00 237.5 F 8.9
24:00 251.8 F 15.7
26:00 263.4 F 22.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 27.8
1:00 271.7 F 28.1
2:00 272.2 F 28.5
3:00 272.6 F 28.6
4:00 272.7 F 28.6

Min 271.4 F 27.7
Max 272.7 F 28.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 234.8 F 0.0
5:00 231.0 F 0.0
10:00 227.6 F 0.0
15:00 224.5 F 0.0
20:00 221.2 F 0.0
25:00 219.7 F 0.0
30:00 217.9 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 27:58
STERILIZING: 4:00
VENTING: 2:47
DRYING: 30:00
TOTAL CYCLE: 01:05:36
CYCLE COMPLETE

12/02/2022 11:14 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____