

-----  
Midmark M11 - v1.0.5  
Total Cycles: 64

Name: RITTER2  
Sterilizer ID: V2313228  
06/01/2021 12:11 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.5 F 0.0  
2:00 150.5 F 2.5  
4:00 169.3 F 6.1  
6:00 195.8 F 11.1  
8:00 212.9 F 5.5  
10:00 226.6 F 10.8  
12:00 231.9 F 8.1  
14:00 240.6 F 11.3  
16:00 246.0 F 13.7  
18:00 258.3 F 20.5  
20:00 269.7 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.2  
1:00 271.9 F 29.3  
2:00 271.8 F 29.2  
3:00 272.1 F 29.1  
4:00 272.2 F 29.3

Min 271.6 F 29.1  
Max 272.6 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.4 F 0.0  
5:00 239.2 F 0.0  
10:00 192.1 F 0.0  
15:00 181.8 F 0.0  
20:00 189.8 F 0.0  
25:00 184.8 F 0.0  
30:00 182.9 F 0.0  
Drying Complete

FILLING: 1:22  
HEATING: 20:50  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:21  
CYCLE COMPLETE

06/01/2021 01:10 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_