
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002764
16:24 30/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:07
PRESSURIZING 3:08
STERILIZING 4:16
136.5 C 226kPa 4:16
136.1 C 220kPa 4:46
136.3 C 223kPa 5:16
136.1 C 219kPa 5:46
136.4 C 226kPa 6:16
136.4 C 225kPa 6:46
136.5 C 225kPa 7:16
136.4 C 222kPa 7:46
136.4 C 222kPa 7:46
VENTING 7:46
AIR DRYING 8:41
CYCLE COMPLETE 8:48

Digital Signature #
4468A68DD1D74277
