

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3940

Name: RITTER 1  
Sterilizer ID: V2178198  
10/24/2023 07:57 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.2 F 0.0  
2:00 71.9 F 0.0  
4:00 88.3 F 0.0  
6:00 126.2 F 0.1  
8:00 157.4 F 0.3  
10:00 180.3 F 3.5  
12:00 202.0 F 6.8  
14:00 210.3 F 10.0  
16:00 224.7 F 12.1  
18:00 229.7 F 7.3  
20:00 240.6 F 11.5  
22:00 234.0 F 7.9  
24:00 245.8 F 12.6  
26:00 243.9 F 11.9  
28:00 256.4 F 18.4  
30:00 267.2 F 25.2  
32:00 272.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.0 F 28.3  
2:00 272.4 F 28.6  
3:00 272.7 F 28.9  
4:00 272.1 F 29.0  
  
Min 271.3 F 28.3  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 170.5 F 0.0  
5:00 219.6 F 0.0  
10:00 183.8 F 0.0  
15:00 176.9 F 0.0  
20:00 199.1 F 0.0  
25:00 202.5 F 0.0  
30:00 185.2 F 0.0  
  
Drying Complete

FILLING: 0:51  
HEATING: 32:10  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:10:19  
CYCLE COMPLETE

10/24/2023 09:07 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_