
Midmark M11 - v1.0.0
Total Cycles: 1447

Name: RITTER 1
Sterilizer ID: V2178198
12/13/2021 07:42 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.4 F 0.0
2:00 72.6 F 0.0
4:00 82.2 F 0.0
6:00 119.5 F 0.0
8:00 173.5 F 0.1
10:00 174.0 F 1.9
12:00 192.1 F 4.3
14:00 213.8 F 7.0
16:00 222.7 F 10.0
18:00 231.6 F 13.5
20:00 231.9 F 7.8
22:00 241.6 F 12.1
24:00 236.0 F 8.8
26:00 242.4 F 10.9
28:00 245.5 F 12.7
30:00 256.8 F 18.7
32:00 266.9 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.7 F 28.7
2:00 271.5 F 28.0
3:00 272.7 F 28.7
4:00 271.7 F 27.9

Min 271.5 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 168.7 F 0.0
5:00 232.8 F 0.0
10:00 225.0 F 0.0
15:00 231.9 F 0.0
20:00 222.0 F 0.0
25:00 220.6 F 0.0
30:00 229.1 F 0.0

Drying Complete

FILLING: 0:57
HEATING: 33:37
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:11:40
CYCLE COMPLETE

12/13/2021 08:53 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____