
Midmark M11 - v1.0.0
Total Cycles: 3531

Name: RITTER 1
Sterilizer ID: V2178198
07/17/2023 07:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 69.6 F 0.0
2:00 72.9 F 0.0
4:00 102.2 F 0.0
6:00 119.8 F 0.1
8:00 146.1 F 1.4
10:00 171.5 F 3.1
12:00 184.2 F 5.4
14:00 196.6 F 8.4
16:00 208.5 F 11.8
18:00 216.0 F 3.9
20:00 226.0 F 7.6
22:00 236.0 F 11.6
24:00 233.4 F 8.1
26:00 245.4 F 13.4
28:00 243.0 F 11.7
30:00 254.2 F 17.4
32:00 264.2 F 23.3
34:00 272.0 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.7 F 28.9
2:00 271.5 F 28.1
3:00 272.0 F 28.7
4:00 272.4 F 28.6

Min 271.2 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.6 F 0.0
5:00 195.8 F 0.0
10:00 192.9 F 0.0
15:00 190.0 F 0.0
20:00 191.3 F 0.0
25:00 193.3 F 0.0
30:00 196.8 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 34:12
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:12:24
CYCLE COMPLETE

07/17/2023 08:49 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____