
Midmark M11 - v1.0.0
Total Cycles: 1997

Name: RITTER 1
Sterilizer ID: V2178198
06/06/2022 03:55 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.3 F 0.0
2:00 174.0 F 3.7
4:00 185.0 F 8.1
6:00 208.8 F 13.2
8:00 225.2 F 8.2
10:00 238.3 F 13.5
12:00 238.6 F 10.5
14:00 234.7 F 8.1
16:00 249.5 F 15.0
18:00 261.1 F 21.6
20:00 271.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.4 F 29.0
2:00 272.2 F 29.0
3:00 272.8 F 29.7
4:00 272.1 F 28.9

Min 271.0 F 28.3
Max 273.4 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.3 F 0.0
5:00 230.4 F 0.0
10:00 210.3 F 0.0
15:00 217.3 F 0.0
20:00 240.7 F 0.0
25:00 233.0 F 0.0
30:00 228.0 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:37
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:14
CYCLE COMPLETE

06/06/2022 04:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____