

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 002469  
16:45 28/01/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:30  
PRESSURIZING 2:32  
STERILIZING 3:35  
136.0 C 220kPa 3:35  
136.4 C 227kPa 4:05  
136.4 C 220kPa 4:35  
136.4 C 227kPa 5:05  
136.4 C 226kPa 5:35  
136.4 C 226kPa 6:05  
136.4 C 227kPa 6:35  
136.4 C 225kPa 7:05  
136.4 C 225kPa 7:05  
VENTING 7:05  
AIR DRYING 8:05  
CYCLE COMPLETE 25:41  
-----

Digital Signature #  
4269A8E23B734E54  
-----