
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 001975
10:20 26/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:56
STERILIZING 4:01
136.0 C 221kPa 4:01
136.3 C 224kPa 4:31
136.3 C 224kPa 5:01
136.4 C 224kPa 5:31
136.0 C 223kPa 6:01
136.2 C 223kPa 6:31
136.2 C 223kPa 7:01
136.1 C 222kPa 7:31
136.1 C 222kPa 7:31
VENTING 7:31
AIR DRYING 8:22
CYCLE COMPLETE 8:38

Digital Signature #
F8FF8F2623189AA0
