

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 003161  
13:23 15/02/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:01  
PRESSURIZING 3:03  
STERILIZING 3:55  
136.4 C 225kPa 3:55  
136.0 C 222kPa 4:25  
136.4 C 224kPa 4:55  
136.3 C 223kPa 5:25  
136.4 C 225kPa 5:55  
136.4 C 225kPa 6:25  
136.4 C 222kPa 6:55  
136.4 C 224kPa 7:25  
136.4 C 224kPa 7:25  
VENTING 7:25  
AIR DRYING 8:17  
CYCLE COMPLETE 8:32

-----  
Digital Signature #  
482547FDBE227CE2  
-----