
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000058
14:59 19/01/2021
WRAPPED
134 C for 6 min.
CYCLE START 0:00
CONDITIONING 4:18
CYCLE ABORTED 15:15
CYCLE FAULT 10
NOT STERILE
END TIME 15:15

Digital Signature #
7E812F35236CD489
