
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006821
10:44 15/01/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:00
PRESSURIZING 3:00
STERILIZING 4:02
136.5 C 223kPa 4:02
136.2 C 222kPa 4:33
136.2 C 222kPa 5:03
136.3 C 224kPa 5:33
136.4 C 222kPa 6:03
136.2 C 220kPa 6:33
136.2 C 221kPa 7:03
136.2 C 217kPa 7:33
136.2 C 217kPa 7:33
VENTING 7:33
AIR DRYING 8:24
CYCLE COMPLETE 9:45

Digital Signature #
3286A0ACC817620B
