
Midmark M11 - v1.0.0
Total Cycles: 2050

Name: RITTER 1
Sterilizer ID: V2178198
06/24/2022 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.1 F 0.0
2:00 164.4 F 1.9
4:00 170.3 F 4.6
6:00 194.5 F 8.2
8:00 218.8 F 13.5
10:00 226.0 F 6.9
12:00 237.4 F 11.5
14:00 233.3 F 8.0
16:00 245.1 F 12.8
18:00 243.2 F 11.8
20:00 255.2 F 18.0
22:00 266.0 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.0
2:00 271.8 F 28.4
3:00 271.4 F 28.5
4:00 272.4 F 29.0

Min 271.1 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 226.1 F 0.0
10:00 238.5 F 0.0
15:00 232.8 F 0.0
20:00 227.9 F 0.0
25:00 227.4 F 0.0
30:00 228.5 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:47
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:01:47
CYCLE COMPLETE

06/24/2022 09:47 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____