

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2802

Name: RITTER 1  
Sterilizer ID: V2178198  
01/23/2023 09:15 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 76.0 F 0.0  
2:00 78.7 F 0.0  
4:00 86.0 F 0.0  
6:00 112.5 F 0.0  
8:00 140.0 F 2.1  
10:00 170.2 F 5.2  
12:00 190.3 F 8.6  
14:00 205.0 F 12.3  
16:00 214.2 F 4.8  
18:00 225.6 F 8.6  
20:00 235.4 F 12.7  
22:00 235.8 F 9.2  
24:00 239.6 F 10.2  
26:00 245.2 F 13.0  
28:00 256.8 F 19.1  
30:00 267.2 F 25.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 29.1  
2:00 271.5 F 28.4  
3:00 272.6 F 29.2  
4:00 271.9 F 28.5

Min 271.4 F 28.3  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 191.9 F 0.0  
5:00 203.0 F 0.0  
10:00 199.5 F 0.0  
15:00 199.1 F 0.0  
20:00 210.5 F 0.0  
25:00 214.5 F 0.0  
30:00 215.0 F 0.0  
Drying Complete

FILLING: 0:47  
HEATING: 31:31  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:09:34  
CYCLE COMPLETE

01/23/2023 10:24 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_