

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2620

Name: RITTER 1  
Sterilizer ID: V2178198  
11/29/2022 01:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 130.9 F 0.0  
2:00 134.7 F 0.8  
4:00 159.2 F 4.0  
6:00 183.4 F 8.0  
8:00 205.3 F 13.1  
10:00 224.2 F 6.9  
12:00 236.0 F 11.8  
14:00 233.8 F 8.5  
16:00 242.1 F 10.9  
18:00 245.3 F 13.0  
20:00 258.0 F 19.7  
22:00 269.5 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.5 F 29.0  
2:00 271.5 F 28.3  
3:00 272.4 F 29.2  
4:00 272.1 F 28.7

Min 271.3 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.2 F 0.0  
5:00 221.3 F 0.0  
10:00 230.7 F 0.0  
15:00 229.0 F 0.0  
20:00 230.8 F 0.0  
25:00 227.1 F 0.0  
30:00 226.9 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 22:60  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:01:06  
CYCLE COMPLETE

11/29/2022 02:46 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_