

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4192

Name: RITTER2  
Sterilizer ID: V2313228  
12/06/2024 10:31 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.8 F 0.0  
2:00 84.6 F 0.0  
4:00 131.2 F 0.0  
6:00 154.2 F 0.2  
8:00 148.9 F 0.5  
10:00 177.4 F 1.3  
12:00 190.1 F 4.3  
14:00 216.6 F 7.5  
16:00 219.8 F 11.2  
18:00 223.3 F 6.0  
20:00 233.7 F 9.5  
22:00 239.1 F 10.9  
24:00 242.7 F 12.1  
26:00 241.0 F 11.0  
28:00 253.3 F 17.2  
30:00 263.9 F 23.5  
32:00 272.1 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 272.7 F 29.3  
2:00 271.6 F 28.4  
3:00 272.7 F 29.4  
4:00 272.1 F 28.7  
  
Min 271.4 F 28.4  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.0 F 0.0  
5:00 229.5 F 0.0  
10:00 205.9 F 0.0  
15:00 175.5 F 0.0  
20:00 170.7 F 0.0  
25:00 168.4 F 0.0  
30:00 167.1 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 32:10  
STERILIZING: 4:00  
VENTING: 3:31  
DRYING: 30:00  
TOTAL CYCLE: 01:10:34  
CYCLE COMPLETE

12/06/2024 11:41 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_