

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4250

Name: RITTER 1  
Sterilizer ID: V2178198  
01/31/2024 08:54 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.6 F 0.0  
2:00 72.3 F 0.0  
4:00 81.9 F 0.7  
6:00 107.6 F 2.7  
8:00 140.9 F 5.5  
10:00 166.4 F 8.8  
12:00 182.8 F 12.4  
14:00 206.9 F 3.4  
16:00 214.6 F 6.9  
18:00 224.2 F 10.3  
20:00 231.0 F 8.8  
22:00 237.0 F 10.1  
24:00 237.1 F 8.9  
26:00 245.6 F 13.0  
28:00 256.7 F 18.8  
30:00 266.6 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.3 F 28.7  
2:00 272.0 F 28.7  
3:00 271.9 F 28.4  
4:00 272.4 F 29.1  
  
Min 271.4 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 176.5 F 0.0  
5:00 225.4 F 0.0  
10:00 215.9 F 0.0  
15:00 223.4 F 0.0  
20:00 224.9 F 0.0  
25:00 226.4 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 0:50  
HEATING: 31:35  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:09:49  
CYCLE COMPLETE

01/31/2024 10:03 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_