
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000566
11:05 06/08/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:05
PRESSURIZING 3:10
STERILIZING 4:22
136.0 C 230kPa 4:22
136.4 C 231kPa 4:52
136.4 C 233kPa 5:22
136.4 C 233kPa 5:52
136.4 C 233kPa 6:22
136.4 C 231kPa 6:52
136.4 C 233kPa 7:22
136.3 C 230kPa 7:52
136.3 C 230kPa 7:52
VENTING 7:52
AIR DRYING 8:46
CYCLE COMPLETE 23:46

Digital Signature #
1FFC197A0B3BBF3D
