
Midmark M11 - v1.0.0
Total Cycles: 2978

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2023 10:38 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.9 F 0.0
2:00 194.9 F 2.8
4:00 219.1 F 7.6
6:00 222.1 F 13.3
8:00 232.0 F 8.7
10:00 237.0 F 10.5
12:00 243.1 F 12.0
14:00 240.9 F 10.6
16:00 254.5 F 17.5
18:00 266.3 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.5 F 28.8
2:00 271.2 F 28.1
3:00 272.7 F 29.0
4:00 271.8 F 28.3

Min 271.2 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 232.0 F 0.0
10:00 229.1 F 0.0
15:00 232.7 F 0.0
20:00 228.1 F 0.0
25:00 226.3 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:36
CYCLE COMPLETE

03/08/2023 11:36 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____