
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000652
9:32 13/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:10
PRESSURIZING 3:16
STERILIZING 4:22
136.1 C 228kPa 4:22
136.4 C 232kPa 4:52
136.1 C 231kPa 5:22
136.3 C 232kPa 5:52
136.1 C 227kPa 6:22
136.3 C 231kPa 6:52
136.0 C 231kPa 7:22
136.2 C 231kPa 7:52
136.2 C 231kPa 7:52
VENTING 7:52
AIR DRYING 8:45
CYCLE COMPLETE 8:53

Digital Signature #
0BD1409DE3FBA43A
