

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 005731  
8:02 12/06/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:28  
PRESSURIZING 3:30  
STERILIZING 4:31  
136.4 C 221kPa 4:31  
136.1 C 222kPa 5:01  
136.0 C 221kPa 5:31  
136.4 C 223kPa 6:01  
136.0 C 219kPa 6:31  
136.4 C 221kPa 7:01  
136.2 C 220kPa 7:31  
136.4 C 223kPa 8:01  
136.4 C 223kPa 8:01  
VENTING 8:02  
AIR DRYING 8:51  
CYCLE COMPLETE 9:20

-----  
Digital Signature #  
C01D155E22CD30DF  
-----