
Midmark M11 - v1.0.0
Total Cycles: 3676

Name: RITTER 1
Sterilizer ID: V2178198
08/18/2023 12:58 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 130.8 F 0.0
2:00 134.3 F 1.3
4:00 143.8 F 3.7
6:00 167.5 F 6.7
8:00 186.2 F 10.1
10:00 204.1 F 9.4
12:00 220.4 F 5.8
14:00 229.7 F 9.7
16:00 236.1 F 10.4
18:00 238.3 F 10.4
20:00 235.9 F 8.4
22:00 247.4 F 14.0
24:00 258.4 F 19.9
26:00 268.3 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.2 F 29.0
2:00 272.4 F 29.4
3:00 271.8 F 28.7
4:00 272.4 F 29.3

Min 271.4 F 28.6
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 238.7 F 0.0
10:00 209.0 F 0.0
15:00 226.0 F 0.0
20:00 216.5 F 0.0
25:00 233.9 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 27:57
STERILIZING: 4:00
VENTING: 3:37
DRYING: 30:00
TOTAL CYCLE: 01:06:25
CYCLE COMPLETE

08/18/2023 02:04 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____