

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5444

Name: RITTER 1  
Sterilizer ID: V2178198  
03/25/2025 03:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 123.8 F 0.0  
2:00 128.9 F 1.9  
4:00 147.6 F 4.8  
6:00 169.5 F 8.1  
8:00 209.9 F 12.0  
10:00 212.6 F 4.2  
12:00 217.1 F 8.5  
14:00 229.5 F 12.9  
16:00 234.6 F 8.8  
18:00 243.0 F 11.6  
20:00 243.6 F 12.1  
22:00 254.8 F 17.9  
24:00 265.0 F 24.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.5  
1:00 271.8 F 29.2  
2:00 272.4 F 29.2  
3:00 272.0 F 29.4  
4:00 272.3 F 29.2

Min 270.9 F 28.3  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 196.7 F 0.0  
10:00 211.3 F 0.0  
15:00 216.4 F 0.0  
20:00 215.8 F 0.0  
25:00 217.6 F 0.0  
30:00 211.0 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 25:53  
STERILIZING: 4:00  
VENTING: 3:55  
DRYING: 30:00  
TOTAL CYCLE: 01:04:43  
CYCLE COMPLETE

03/25/2025 04:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_