

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 001087  
10:58 20/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:45  
STERILIZING 3:51  
136.3 C 220kPa 3:51  
136.0 C 222kPa 4:21  
136.4 C 224kPa 4:51  
136.2 C 224kPa 5:21  
136.4 C 222kPa 5:51  
136.3 C 223kPa 6:21  
136.3 C 224kPa 6:51  
136.4 C 224kPa 7:21  
136.4 C 224kPa 7:21  
VENTING 7:21  
AIR DRYING 8:12  
CYCLE COMPLETE 23:13  
-----

Digital Signature #  
BAC7265DA52FAA9C  
-----