
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008373
16:49 27/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:00
PRESSURIZING 2:02
STERILIZING 3:01
136.4 C 221kPa 3:01
136.2 C 223kPa 3:31
136.3 C 222kPa 4:01
136.2 C 224kPa 4:31
136.3 C 224kPa 5:01
136.3 C 222kPa 5:31
136.4 C 224kPa 6:01
136.4 C 224kPa 6:31
136.4 C 224kPa 6:31
VENTING 6:31
AIR DRYING 7:20
CYCLE COMPLETE 26:29

Digital Signature #
2676245F0DAFE901
