

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1020

Name: RITTER 1  
Sterilizer ID: V2178198  
07/15/2021 01:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 200.1 F 0.0  
2:00 187.0 F 3.6  
4:00 213.5 F 8.2  
6:00 221.8 F 10.6  
8:00 232.4 F 8.9  
10:00 238.6 F 10.6  
12:00 243.8 F 12.3  
14:00 241.9 F 11.1  
16:00 255.3 F 17.9  
18:00 266.6 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 272.4 F 28.6  
2:00 271.2 F 28.1  
3:00 272.5 F 28.7  
4:00 271.4 F 28.0  
  
Min 271.1 F 27.9  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.3 F 0.0  
5:00 231.5 F 0.0  
10:00 226.5 F 0.0  
15:00 227.1 F 0.0  
20:00 225.8 F 0.0  
Dry Cycle Aborted

FILLING: 2:00  
HEATING: 19:32  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 23:42  
TOTAL CYCLE: 00:52:26

07/15/2021 01:58 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02