
Midmark M11 - v1.0.5
Total Cycles: 1296

Name: RITTER2
Sterilizer ID: V2313228
07/14/2022 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.2 F 0.0
2:00 151.5 F 0.1
4:00 181.6 F 2.4
6:00 201.6 F 7.1
8:00 221.9 F 12.4
10:00 228.5 F 8.1
12:00 239.8 F 12.8
14:00 241.0 F 11.8
16:00 239.9 F 10.8
18:00 253.4 F 17.6
20:00 264.9 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.3 F 29.3
2:00 271.9 F 29.6
3:00 272.0 F 29.1
4:00 272.0 F 29.7

Min 271.4 F 28.8
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.2 F 0.0
5:00 217.2 F 0.0
10:00 224.1 F 0.0
15:00 221.1 F 0.0
20:00 192.4 F 0.0
25:00 184.7 F 0.0
30:00 191.6 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 21:50
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:11
CYCLE COMPLETE

07/14/2022 10:12 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____