
Midmark M11 - v1.0.5
Total Cycles: 2681

Name: RITTER2
Sterilizer ID: V2313228
07/21/2023 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.4 F 0.0
2:00 136.3 F 0.0
4:00 141.9 F 0.0
6:00 187.4 F 0.0
8:00 210.4 F 0.1
10:00 216.5 F 2.5
12:00 227.6 F 5.5
14:00 236.1 F 9.0
16:00 245.1 F 13.0
18:00 241.4 F 11.2
20:00 238.5 F 10.0
22:00 238.9 F 9.6
24:00 250.3 F 15.6
26:00 261.3 F 21.9
28:00 271.4 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 272.7 F 29.2
2:00 271.7 F 28.7
3:00 272.7 F 29.2
4:00 271.8 F 28.5

Min 271.6 F 28.4
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.3 F 0.0
5:00 218.0 F 0.0
10:00 234.2 F 0.0
15:00 227.1 F 0.0
20:00 225.0 F 0.0
25:00 227.8 F 0.0
30:00 229.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 28:40
STERILIZING: 4:00
VENTING: 3:26
DRYING: 30:00
TOTAL CYCLE: 01:07:01
CYCLE COMPLETE

07/21/2023 03:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____