
Midmark M11 - v1.0.5
Total Cycles: 3296

Name: RITTER2
Sterilizer ID: V2313228
01/12/2024 05:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.7 F 0.0
2:00 144.5 F 2.3
4:00 169.2 F 6.7
6:00 197.0 F 12.6
8:00 219.1 F 6.5
10:00 234.8 F 12.3
12:00 237.0 F 10.0
14:00 235.7 F 8.6
16:00 252.0 F 16.5
18:00 265.3 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.5 F 28.9
2:00 272.6 F 29.4
3:00 271.8 F 28.5
4:00 273.0 F 29.4

Min 271.6 F 28.3
Max 273.0 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 170.4 F 0.0
5:00 196.5 F 0.0
10:00 216.3 F 0.0
15:00 216.3 F 0.0
20:00 212.3 F 0.0
25:00 214.1 F 0.0
30:00 214.8 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 19:35
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

01/12/2024 06:03 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____