
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001878
10:34 22/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:23
STERILIZING 3:27
136.0 C 221kPa 3:27
136.1 C 223kPa 3:57
136.3 C 225kPa 4:27
136.4 C 226kPa 4:57
136.4 C 223kPa 5:27
136.2 C 224kPa 5:57
136.4 C 225kPa 6:27
136.4 C 223kPa 6:57
136.4 C 223kPa 6:57
VENTING 6:57
AIR DRYING 7:53
CYCLE COMPLETE 8:30

Digital Signature #
B450DC78F7FFFC1
